



**Agilent Technologies**

**DECLARATION OF CONFORMITY**  
According to EN ISO/IEC 17050-1:2004



**Manufacturer's Name:** Agilent Technologies International Japan, Ltd.  
**Manufacturer's Address:** 9-1, Takakura-machi, Hachioji-shi, Tokyo, 192-0033 Japan

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Inductively Coupled Plasma Mass Spectrometer  
Agilent 7800 ICP-MS and 7900 ICP-MS  
**Model Number:** G8421A and G8403A  
**Product Options:** This declaration covers all options of the above products.  
**Serial Number:** JPXXXXXXXX

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

EMC Directive 2014/30/EU  
Machinery Directive 2006/42/EC  
RoHS Directive 2011/65/EU

**and conforms with the following product standards:**

EMC	Standard	Limit
	IEC 61326-1:2012 / EN 61326-1:2013	
	CISPR 11:2009+A1:2010 / EN 55011:2009+A1:2010	Group 1 Class A
	IEC 61000-4-2:2008 / EN 61000-4-2:2009	4 kV CD, 8 kV AD
	IEC 61000-4-3:2006+A1 :2007+A2 :2010	3 V/m, 80-1000 MHz
	/ EN 61000-4-3:2006+A1 :2008+A2 :2010	3 V/m, 1.4-2 GHz
	IEC 61000-4-4:2004+corr. Jun :2007+A1:2010	1 V/m, 2-2.7 GHz
	/ EN 61000-4-4:2004+A1 :2010	0.5 kV signal lines, 1 kV power lines
	IEC 61000-4-5:2005+corr.October / EN 61000-4-5:2006	0.5 kV line-line, 1 kV line-ground
	IEC 61000-4-6:2008 / EN 61000-4-6:2009	3 V, 0.15-80 MHz
	Canada: ICES-001:2006	

The product was tested in a typical configuration with Agilent Technologies test systems.

Safety	Standard
	IEC 61010-1:2010 / EN 61010-1:2010
	IEC 61010-2-061:2015 / EN 61010-2-061:2015
	IEC 61010-2-081:2015 / EN 61010-2-081:2015
	Canada: CAN/CSA C22.2 No. 61010-1-12
	Canada: CAN/CSA C22.2 No. 61010-2-061:15
	Canada: CAN/CSA C22.2 No. 61010-2-081:15
	USA: UL 61010-1 (3rd Edition)
	USA: UL 61010-2-061:2015
	USA: UL 61010-2-081:2015
	IEC 62311:2007 / EN 62311:2008

Other	Standard
	EN 50581:2012

**Contact established in the Community authorized to compile the technical file or the relevant technical documents:**

Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany

**Supplementary Information:**

This ISM device complies with Canadian ICES-001.  
Cet appareil ISM est conforme à la norme NMB-001 du Canada

**This DoC applies to above-listed products placed on the EU market after:**

May 16, 2017

Date

Shunichiro Cho  
Quality Manager

For more information, please contact your local Agilent Technologies sales office, agent or distributor.